

SPONSOR 12364

REPORT DATE: 8/23/12

Richard Balcomb
Intertek on behalf of: HaloPolymer
1035 17th Street No.4
Santa Monica, CA 90403



MSP NUMBER: 12D0300H

P.O. NUMBER: USA01-0000207606

PAGE: 1 of 1

CERTIFICATE

CLASS IV – 50°C PLASTIC TESTS (USP 35 <88>, REV. 5/2012)

Name: HaloPolymer Kirovo-Chepetsk, Fluorplast-4D
Physical Description: Solid
Total Quantity Received for Testing: 1 bag containing 4 pieces of each lot
Lot Number: 62054/62055 (Implant test);
62572/62573 (Intracutaneous and Systemic test)
Storage Condition: Room Temperature
Date Received: 4/18/12 (lot 62054/62055)
7/6/12 (lot 62572/62573)

Class IV - Intracutaneous Test
PBL Report No.: 12D0300H-X02G

Class IV - Systemic Injection Test
PBL Report No.: 12D0300H-X03G

Class IV - Implantation Test
PBL Report No.: 12D0300H-X04G


The test article met the requirements for Plastic Class IV - 50°C testing per USP <88>.



Multiple Study Project Coordinator
Pacific BioLabs

8/23/12

Date



QAU Review

8/23/12

Date

:wc