

SPONSOR 12364

Richard Balcomb
Intertek on behalf of: HaloPolymer
1035 17th Street No.4
Santa Monica, CA 90403

REPORT DATE: 7/23/12



MSP NUMBER: 12D0298H

P.O. NUMBER: USA01-0000207606

PAGE: 1 of 1

CERTIFICATE

CLASS IV – 50°C PLASTIC TESTS (USP 35 <88>, REV. 5/2012)


Name: HaloPolymer Kirovo-Chepetsk, Fluorplast-4
Physical Description: Solid
Total Quantity Received for Testing: 1 bag containing 4 pieces, 1 bag containing 3 pieces
Lot Number: Series 945-1271
Storage Condition: Room Temperature
Date Received: 4/18/12, 6/1/12

Class IV - Systemic Injection Test
PBL Report No.: 12D0298H-X03G

Class IV - Intracutaneous Test
PBL Report No.: 12D0298H-X02G


Class IV - Implantation Test
PBL Report No.: 12D0298H-X04G

The test article met the requirements for Plastic Class IV - 50°C testing per USP <88>.



Multiple Study Project Coordinator
Pacific BioLabs

7/23/12
Date



QAU Review

7/23/12
Date