

SPONSOR 12364

Richard Balcomb
Intertek on behalf of: HaloPolymer
1035 17th Street No.4
Santa Monica, CA 90403

REPORT DATE: 7/18/12



MSP NUMBER: 12E0278H

P.O. NUMBER: USA01-0000207606

PAGE: 1 of 1

CERTIFICATE

CLASS IV – 50°C PLASTIC TESTS (USP 35 <88>, REV. 5/2012)


Name: HaloPolymer Perm, Fluorplast-4D
Physical Description: plastic sheet
Total Quantity Received for Testing: 1 bag containing 10 plaques
Lot Number: Batch 753
Storage Condition: Room Temperature
Date Received: 5/17/12

Class IV.- Intracutaneous Test
PBL Report No.: 12E0278H-X02G

Class IV - Systemic Injection Test
PBL Report No.: 12E0278H-X03G

Class IV - Implantation Test
PBL Report No.: 12E0278H-X04G

The test article met the requirements for Plastic Class IV - 50°C testing per USP <88>.



Multiple Study Project Coordinator
Pacific BioLabs

7/19/12

Date

Ma. Melanie Pacheco

QAU Review

07/19/12

Date